



CERTIFICATE OF ANALYSIS

Column Regenerant for Sodium-based Amino Acids Analysis

RG011

Product: Sodium Column Regenerant

Part Number: RG011

Lot Number: 112003

Production Date: 05 Dec 2011

Expiration Date: 05 Dec 2013

Primary Application: For column regeneration in sodium-based amino acids analysis

Acceptance Criteria:

1. pH = 12.74 ± 0.10 @ 25 ± 0.4 °C (Y/N)
2. Conductivity meets specifications (Y/N)

Y
Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

John P. [Signature]
VP QA
Title

Date:

28 Feb 2012

Rev. 11, 11/10/2011

ACCEPTED

PICKERING LABORATORIES, INC
QUALITY ASSURANCE

P:\Documentation\Chemical Documentation\RG011