



CERTIFICATE OF ANALYSIS

Column Regenerant for Sodium-based Amino Acids Analysis

RG011

Product: Sodium Column Regenerant

Part Number: RG011

Lot Number: 2101018

Production Date: January 13, 2021

Expiration Date: January 13, 2023

Primary Application: For column regeneration in sodium-based amino acids analysis

Acceptance Criteria:

1. pH = 12.74 ± 0.10 @ 25 ± 0.4 °C (Y/N)
2. Conductivity meets specifications (Y/N)

Y

Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy
President

Title

Date: January 28, 2021