



CERTIFICATE OF ANALYSIS

Column Regenerant for Sodium-based Amino Acids Analysis

RG011

Product: Sodium Column Regenerant

Part Number: RG011

Lot Number: 509006

Production Date: 04 Sep 2015

Expiration Date: 04 Sep 2017

Primary Application: For column regeneration in sodium-based amino acids analysis

Acceptance Criteria:

1. pH = 12.74 ± 0.10 @ 25 ± 0.4 °C (Y/N)

y

2. Conductivity meets specifications (Y/N)

y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy

Date: September 11, 2015

President

Title

