



## CERTIFICATE OF ANALYSIS

### Column Regenerant for Sodium-based Amino Acids Analysis

**RG011**

**Product:** Sodium Column Regenerant

**Part Number:** RG011

**Lot Number:** 604024

**Production Date:** 21 Apr 2016

**Expiration Date:** 21 Apr 2018

**Primary Application:** For column regeneration in sodium-based amino acids analysis

**Acceptance Criteria:**

- |  |          |
|--|----------|
| 1. pH = $12.74 \pm 0.10$ @ $25 \pm 0.4$ °C (Y/N) | <u>y</u> |
| 2. Conductivity meets specifications (Y/N)       | <u>y</u> |

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** April 25, 2016  
President  
Title