



## CERTIFICATE OF ANALYSIS

### Column Regenerant for Sodium-based Amino Acids Analysis

RG011

**Product:** Sodium Column Regenerant

**Part Number:** RG011

**Lot Number:** 903046

**Production Date:** 26 Mar 2019

**Expiration Date:** 26 Mar 2021

**Primary Application:** For column regeneration in sodium-based amino acids analysis

**Acceptance Criteria:**

- 1. pH =  $12.74 \pm 0.10$  @  $25 \pm 0.4$  °C (Y/N) y
- 2. Conductivity meets specifications (Y/N) y

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** March 28, 2019  
President  
Title