



## CERTIFICATE OF ANALYSIS

### Column Regenerant for Glyphosate Ion-exchange (K<sup>+</sup>) Column RG019

**Product:** Column Regenerant for Glyphosate and AMPA Analysis

**Part Number:** RG019

**Lot Number:** 004018

**Production Date:** April 14, 2020

**Expiration Date:** April 14, 2022

**Primary Application:** Column Regenerant for ion-exchange chromatography of Glyphosate and AMPA.

**Acceptance Criteria:**

- |  |          |
|--|----------|
| 1. pH = $11.75 \pm 0.24$ @ $25 \pm 0.4$ °C (Y/N) | <u>Y</u> |
| 2. Conductivity = $1344 \pm 90$ uS/cm (Y/N)      | <u>Y</u> |

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:** Rebecca Smith **Date:** April 24, 2020  
Vice President  
Title