



CERTIFICATE OF ANALYSIS

Column Regenerant for Glyphosate and AMPA Systems

RG019

Product: RG019 Regenerant for Glyphosate and AMPA Systems

Part Number: RG019

Lot Number: 006033

Test Start Date: 21 Jun 2010

Expiration Date: 21 Jun 2012

Primary Application: Regenerant for ion-exchange chromatography of Glyphosate and AMPA.


Acceptance Criteria:

- 1. pH = 11.75 ± 0.24 @ 25 ± 0.4 °C (Y/N)
- 2. Conductivity = 1344 ± 90 uS/cm (Y/N)

Actual Values

11.525
1356

Storage Conditions: Dry at 8 to 25

Analysis Certified by:  Date: 23 Jun 2010
VP Ops
Title

Rev. 8, 8/11/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE