



CERTIFICATE OF ANALYSIS

Column Regenerant for Glyphosate and AMPA Systems

RG019

Product: RG019 Regenerant for Glyphosate and AMPA Systems

Part Number: RG019

Lot Number: 008013

Test Start Date: 11 Aug 2010

Expiration Date: 11 Aug 2012

Primary Application: Regenerant for ion-exchange chromatography of Glyphosate and AMPA.

Acceptance Criteria:


1. pH = 11.75 ± 0.24 @ 25 ± 0.4 °C (Y/N)
2. Conductivity = 1344 ± 90 uS/cm (Y/N)

Actual Values

11.588
1372

Storage Conditions: Dry at 8 to 25

Analysis Certified by: _____


VP Operator

Title

Date: _____

23 Aug 2010

Rev. 8, 8/11/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE