



# CERTIFICATE OF ANALYSIS

## Column Regenerant for Glyphosate and AMPA Systems

### RG019

**Product:** RG019 Regenerant for Glyphosate and AMPA Systems

**Part Number:** RG019

**Lot Number:** 011004

**Test Start Date:** 02 Nov 2010

**Expiration Date:** 02 Nov 2012

**Primary Application:** Regenerant for ion-exchange chromatography of Glyphosate and AMPA.

**Acceptance Criteria:**

1. pH =  $11.75 \pm 0.24$  @  $25 \pm 0.4$  °C (Y/N)
2. Conductivity =  $1344 \pm 90$  uS/cm (Y/N)

**Actual Values**

11.652

1346

**Storage Conditions:** Dry at 8 to 25

**Analysis Certified by:** \_\_\_\_\_

*[Signature]*  
VP Ops

Title

**Date:** 08 Nov 2010

Rev. 8, 8/11/08

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE