



CERTIFICATE OF ANALYSIS

Column Regenerant for Glyphosate and AMPA Systems

RG019

Product: RG019 Regenerant for Glyphosate and AMPA Systems

Part Number: RG019

Lot Number: 101033

Test Start Date: 25 JAN 2011

Expiration Date: 25 JAN 2013

Primary Application: Regenerant for ion-exchange chromatography of Glyphosate and AMPA.

Acceptance Criteria:

1. pH = 11.75 ± 0.24 @ 25 ± 0.4 °C (Y/N)
2. Conductivity = 1344 ± 90 uS/cm (Y/N)

Actual Values

11.630
1377

Storage Conditions: Dry at 8 to 25

Analysis Certified by: _____

Jim Murphy
VP OPR

Title

Date: 15 FEB 2011

Rev. 8, 8/11/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE