



CERTIFICATE OF ANALYSIS

Column Regenerant for Glyphosate and AMPA Analysis RG019

Product: RG019 Column Regenerant for Glyphosate and AMPA Analysis

Part Number: RG019

Lot Number: 103043

Production Date: 28 MAR 2011

Expiration Date: 28 MAR 2013

Primary Application: Column Regenerant for ion-exchange chromatography of Glyphosate and AMPA.

Acceptance Criteria:

1. pH = 11.75 ± 0.24 @ 25 ± 0.4 °C (Y/N)
2. Conductivity = 1344 ± 90 uS/cm (Y/N)

Y
Y

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by:

Jim Murphy

Date: 30 MAR 2011

VP Ops
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 9, 04/08/2011