



CERTIFICATE OF ANALYSIS

Column Regenerant for Ion-exchange Chromatography of Glyphosate and AMPA Analysis RG019

Product: Column Regenerant for Glyphosate and AMPA Analysis
Part Number: RG019
Lot Number: 109011
Production Date: 16 Sep 2011
Expiration Date: 16 Sep 2013

Primary Application: Column Regenerant for ion-exchange chromatography of
Glyphosate and AMPA.

Acceptance Criteria:

1. pH = 11.75 ± 0.24 @ 25 ± 0.4 °C (Y/N)
2. Conductivity = 1344 ± 90 uS/cm (Y/N)

4
14

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: _____

[Signature]
VP OPR
Title

Date: 22 Sep 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE