



## CERTIFICATE OF ANALYSIS

### Column Regenerant for Ion-exchange Chromatography of Glyphosate and AMPA Analysis RG019

**Product:** Column Regenerant for Glyphosate and AMPA Analysis

**Part Number:** RG019

**Lot Number:**

112017

**Production Date:**

13 Dec 2011

**Expiration Date:**

12 Dec 2013

**Primary Application:** Column Regenerant for ion-exchange chromatography of  
Glyphosate and AMPA.

**Acceptance Criteria:**

1. pH =  $11.75 \pm 0.24$  @  $25 \pm 0.4$  °C (Y/N)
2. Conductivity =  $1344 \pm 90$  uS/cm (Y/N)

Y  
Y

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:**

*[Signature]*  
VIP OPER  
Title

**Date:**

28 Feb 2012

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE