



CERTIFICATE OF ANALYSIS

Column Regenerant for Glyphosate Ion-exchange (K⁺) Column RG019

Product: Column Regenerant for Glyphosate and AMPA Analysis

Part Number: RG019

Lot Number: 208030

Production Date: 29 Aug 2012

Expiration Date: 29 Aug 2014

Primary Application: Column Regenerant for ion-exchange chromatography of Glyphosate and AMPA.

Acceptance Criteria:

1. pH = 11.75 ± 0.24 @ 25 ± 0.4 °C (Y/N)

y

2. Conductivity = 1344 ± 90 uS/cm (Y/N)

y

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: Jim Murphy

Date: 12 Sep 2012

VP Operations

Title