



## CERTIFICATE OF ANALYSIS

### Column Regenerant for Glyphosate Ion-exchange (K<sup>+</sup>) Column RG019

**Product:** Column Regenerant for Glyphosate and AMPA Analysis

**Part Number:** RG019

**Lot Number:** 2110052

**Production Date:** October 22, 2021

**Expiration Date:** October 22, 2023

**Primary Application:** Column Regenerant for ion-exchange chromatography of Glyphosate and AMPA.

**Acceptance Criteria:**

1. pH =  $11.75 \pm 0.24$  @  $25 \pm 0.4$  °C (Y/N)
2. Conductivity =  $1344 \pm 90$  uS/cm (Y/N)

Y

Y

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:** Jim Murphy  
President

Title

**Date:** October 26, 2021