



CERTIFICATE OF ANALYSIS

Column Regenerant for Glyphosate Ion-exchange (K⁺) Column RG019

Product: Column Regenerant for Glyphosate and AMPA Analysis

Part Number: RG019

Lot Number: 807036

Production Date: 25 Jul 2018

Expiration Date: 25 Jul 2020

Primary Application: Column Regenerant for ion-exchange chromatography of Glyphosate and AMPA.

Acceptance Criteria:

1. pH = 11.75 ± 0.24 @ 25 ± 0.4 °C (Y/N) y
2. Conductivity = 1344 ± 90 uS/cm (Y/N) y

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: Jim Murphy **Date:** July 31, 2018
President
Title