



CERTIFICATE OF ANALYSIS

Regenerant for Sodium System RG011 / 1700-1120

Product: Regenerant for Sodium System

Part Number: RG011 / 1700-1120

Lot Number: 911016

Test Date: 11 Nov 09

Expiration Date: 13 Nov 10

Primary Application: Sodium-based amino acid analysis

Acceptance Criteria:

Actual Values:

1. pH = 12.74 ± 0.10 @ 25 ± 0.4 °C
2. Conductivity = 31.3 ± 0.7 @ 25 ± 0.4 °C

12.678
31.1

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jan Murphy

Date: _____

19 Nov 09

VP Operations

Title

Rev. 9, 7/09/08

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE