



CERTIFICATE OF ANALYSIS

SERAPREP

SP100 (1800-0020)

Product: SERAPREP
Part Number: SP100 (1800-0020)
Lot Number: 009015
Test Start Date: 20 Sep 2010

Primary Application: Prepare serum and other native samples for amino acid analysis.

Acceptance Criteria:

1. pH = 1.04 ± 0.02 @ 25 °C

Actual Values

1.060

2. Use threonine and serine peaks for evaluation. Serum sample peak times ± 30 seconds of calibration standard. (Y/N)

Y

Calibration Standard:

Physiological Fluid Calibration Standard, PN 011006p, LN 909011

Analysis Certified by: *Jim Murphy* Date: 22 Sep 2010

V.P. Ope

Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE