



CERTIFICATE OF ANALYSIS

SERAPREP

SP100 (1800-0020)

Product: SERAPREP
Part Number: SP100 (1800-0020)
Lot Number: 012012
Test Start Date: 08 Dec 2010

Primary Application: Prepare serum and other native samples for amino acid analysis.

Acceptance Criteria:

- | | Actual Values |
|---|----------------------|
| 1. pH = 1.04 ± 0.02 @ 25 °C | <u>1.050</u> |
| 2. Use threonine and serine peaks for evaluation. Serum sample peak times ± 30 seconds of calibration standard. (Y/N) | <u>Y</u> |

Calibration Standard:

Physiological Fluid Calibration Standard, PN 011006p, LN 909011

Analysis Certified by:  Date: 14 Dec 2010

VP Opr
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE