



CERTIFICATE OF ANALYSIS

SERAPREP

SP100 (1800-0020)

Product: SERAPREP
Part Number: SP100 (1800-0020)
Lot Number: 108003
Production Date: 03 Aug 2011
Expiration Date: 02 Aug 2012

Primary Application: Prepare serum and other native samples for amino acid analysis.

Acceptance Criteria:

1. pH = 1.04 ± 0.02 @ 25 °C (Y/N) 4
2. Use threonine and serine peaks for evaluation. Serum sample peak times ± 30 seconds of calibration standard. (Y/N) 4

Calibration Standard:

Physiological Fluid Calibration Standard, PN 011006P

Analysis Certified by:

[Signature]
VP OPR
Title

Date:

20 Sep 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE