

## **CERTIFICATE OF ANALYSIS**

## SERAPREP


## SP100

Product: SERAPREP	
Part Number: SP100	
Lot Number:	312017
Production Date:	16 Dec 2013
Expiration Date:	16 Dec 2014

**Primary Application:** Preparation of serum and other native samples for amino acids analysis

## Acceptance Criteria:

- pH is 1.04 ± 0.02 at 25° C ± 0.4°C (Y/N)
  Retention times difference for Threonine and Serine meets specification (Y/N)
- 3. Baseline: no peaks (Y/N)



Sample: Serum sample prepared with SERAPREP

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:	Jim Murphy	Date:	24 Jan 2014	
	President			
	Title	-		

Rev. A, 11/09/12

P:\Documentation\Chemical Documentation\SP100