

## **CERTIFICATE OF ANALYSIS**

## SERAPREP


## SP100

<b>Product: SERAPREP</b>	
Part Number: SP100	
Lot Number:	902024
Production Date:	13 Feb 2019
Expiration Date:	13 Feb 2020

**Primary Application:** Preparation of serum and other native samples for amino acids analysis

## Acceptance Criteria:

- 1. pH is 1.04  $\pm$  0.02 at 25° C  $\pm$  0.4°C (Y/N)
- 2. Retention times difference for Threonine and Serine meets specification (Y/N)
- 3. Baseline: no peaks (Y/N)



Sample: Serum sample prepared with SERAPREP

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:	Jim Murphy	Date:	February 25, 2019
	President		
	Title		

Rev. A, 11/09/12

P:\Documentation\Chemical Documentation\SP100