

CERTIFICATE OF ANALYSIS

TRIONE, Ninhydrin Reagent T100c

Product: TRIONE, Ninhydrin Reagent

Part Number: T100c

Lot Number: 009016

Test Start Date: 09 Sep 2010

Expiration Date: 12 Sep 09 DEC 2010 *(Handwritten initials)*

Primary Application: Amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple (Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.3

Actual Values:

4
4

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

(Handwritten signature)
VP O&R

Title

Date: _____

17 Sep 2010

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE