



CERTIFICATE OF ANALYSIS

TRIONE, Ninhydrin Reagent T100c

Product: TRIONE, Ninhydrin Reagent

Part Number: T100c

Lot Number: 103028

Production Date: 21 MAR 2011

Expiration Date: 18 JUN 2011

Primary Application: Amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple (Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.3

4
4

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim [Signature]
VP OPR

Date: 23 MAR 2011

Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 8, 3/25/2011