



CERTIFICATE OF ANALYSIS

TRIONE, Ninhydrin Reagent T100c

Product: TRIONE, Ninhydrin Reagent

Part Number: T100c

Lot Number: 109004

Production Date: 03 Sep 2011

Expiration Date: 02 Dec 2011

Primary Application: Amino acid analysis


Acceptance Criteria:

1. TRIONE color without shades of brown or purple (Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.3 (Y/N)

 4
 4

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____



VP OPR

Date: 20 Sep 2011

Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 8, 3/25/2011