



CERTIFICATE OF ANALYSIS

TRIONE, Ninhydrin Reagent T100/T100C

Product: TRIONE, Ninhydrin Reagent

Part Number: T100/T100C

Lot Number: 109042

Production Date: 03 Oct 2011

Expiration Date: 29 Dec 2011

Primary Application: Amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple (Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.0 (Y/N)

Y
Y

Storage Conditions: Dry at 8 to 25 °C

Sample: Aspartic acid and Proline mixture.

Analysis Certified by: *Jim Murphy*

Date: 05 Oct 2011

JF Ope

Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE