



CERTIFICATE OF ANALYSIS

TRIONE Ninhydrin Reagent T100/T100C

Product: TRIONE, Ninhydrin Reagent

Part Number: T100/T100C

Lot Number: 501008

Production Date: 07 Jan 2015

Expiration Date: 07 May 2015

Primary Application: Amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple (Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.0 (Y/N)

y
y

Storage Conditions: Dry at 8 to 25 °C

Sample: Aspartic acid and Proline mixture.

Analysis Certified by: Jim Murphy
President
Title

Date: 13 Jan 2015