



CERTIFICATE OF ANALYSIS
TRIONE, Ninhydrin Reagent (2 Parts)
T200

Product: TRIONE, Ninhydrin Reagent (2 Parts)

Part Number: T200

Lot Number: 008027

Test Start Date: 19 AUG 2010

Expiration Date: 18 AUG 2011

Primary Application: Sodium amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple.
2. L-aspartic acid to L-proline ratio ≥ 3.3

Actual Values:

4
4

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Rumpf
VP OPR
Title

Date: _____

30 Aug 2010

Rev. 6, 8/11/08

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE