



## CERTIFICATE OF ANALYSIS

### TRIONE, Ninhydrin Reagent (2 Parts) T200

Product: TRIONE, Ninhydrin Reagent (2 Parts)

Part Number: T200

Lot Number: 111033

Production Date: 28 Nov 2011

Expiration Date: 22 Nov 2012

Primary Application: Amino acid analysis

#### Acceptance Criteria:

1. TRIONE color without shades of brown or purple(Y/N)
2. L-aspartic acid to L-proline ratio  $\geq 3.0$  (Y/N)

Y  
Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: \_\_\_\_\_

*Jim P. [Signature]*  
VP Opr

Title

Date: 09 Feb 2012

# ACCEPTED

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE