



CERTIFICATE OF ANALYSIS

TRIONE Ninhydrin Reagent (2 Parts) T200

Product: TRIONE, Ninhydrin Reagent (2 Parts)

Part Number: T200

Lot Number: 202006

Production Date: 07 Feb 2012

Expiration Date: 22 Feb 2013

Primary Application: Amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple (Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.0 (Y/N)

Y
Y

Storage Conditions: Dry at 8 to 25 °C

Sample: Aspartic acid and Proline mixture.

Analysis Certified by:

[Signature]
VT OPR
Title

Date:

28 Feb 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE