



CERTIFICATE OF ANALYSIS

TRIONE Ninhydrin Reagent (2 Parts) T200

Product: TRIONE, Ninhydrin Reagent (2 Parts)

Part Number: T200

Lot Number: 308035

Production Date: 29 Aug 2013

Expiration Date: 29 Aug 2014

Primary Application: Amino acid analysis

Acceptance Criteria:

- | | |
|---|----------|
| 1. TRIONE color without shades of brown or purple (Y/N) | <u>y</u> |
| 2. L-aspartic acid to L-proline ratio ≥ 3.0 (Y/N) | <u>y</u> |

Storage Conditions: Dry at 8 to 25 °C

Sample: Aspartic acid and Proline mixture.

Analysis Certified by: Jim Murphy **Date:** 13 Sep 2013
President
Title