



# CERTIFICATE OF ANALYSIS

**URIPREP**

**UP100**

**Product:** URIPREP  
**Part Number:** UP100  
**Lot Number:** 004016  
**Production Date:** April 13, 2020  
**Expiration Date:** April 13, 2021

**Primary Application:** Preparation of urine and other native samples for amino acids analysis

**Acceptance Criteria:**

- |  |          |
|--|----------|
| 1. pH is $1.29 \pm 0.01$ at $25^{\circ} \text{C} \pm 0.4^{\circ} \text{C}$ (Y/N) | <u>Y</u> |
| 2. Retention times difference for Threonine and Serine meets specification (Y/N) | <u>Y</u> |
| 3. Baseline: no peaks (Y/N)  | <u>Y</u> |

**Sample:** Urine sample prepared with URIPREP

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Rebecca Smith **Date:** April 29, 2020  
Vice President  
Title