



# CERTIFICATE OF ANALYSIS

**URIPREP**

**UP100**

**Product:** URIPREP  
**Part Number:** UP100  
**Lot Number:** 601029  
**Production Date:** 22 Jan 2016  
**Expiration Date:** 22 Jan 2017

**Primary Application:** Preparation of urine and other native samples for amino acids analysis

**Acceptance Criteria:**

- 1. pH is  $1.29 \pm 0.01$  at  $25^{\circ}\text{C} \pm 0.4^{\circ}\text{C}$  (Y/N) y
- 2. Retention times difference for Threonine and Serine meets specification (Y/N) y
- 3. Baseline: no peaks (Y/N) y

**Sample:** Urine sample prepared with URIPREP

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** January 29, 2016  
President  
Title