



CERTIFICATE OF ANALYSIS

URIPREP

UP100

Product: URIPREP
Part Number: UP100
Lot Number: 610024
Production Date: 19 Oct 2016
Expiration Date: 19 Oct 2017

Primary Application: Preparation of urine and other native samples for amino acids analysis

Acceptance Criteria:

- 1. pH is 1.29 ± 0.01 at $25^{\circ}\text{C} \pm 0.4^{\circ}\text{C}$ (Y/N) y
- 2. Retention times difference for Threonine and Serine meets specification (Y/N) y
- 3. Baseline: no peaks (Y/N) y

Sample: Urine sample prepared with URIPREP

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** October 25, 2016
President
Title