

## **CERTIFICATE OF ANALYSIS**

## URIPREP


## UP100

Product: URIPREP	
Part Number: UP100	
Lot Number:	812017
Production Date:	13 Dec 2018
Expiration Date:	13 Dec 2019

**Primary Application:** Preparation of urine and other native samples for amino acids analysis

## Acceptance Criteria:

- pH is 1.29 ± 0.01 at 25° C ± 0.4°C (Y/N)
  Retention times difference for Threonine and Serine meets specification (Y/N)
- 3. Baseline: no peaks (Y/N)



Sample: Urine sample prepared with URIPREP

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:	Jim Murphy	Date:	December 17,2018
	Title	-	

Rev. A, 11/09/12

P:\Documentation\Chemical Documentation\UP100